

ALTAIR5/5X Multigas Detector Quick Start Card

Please review the back of this card for a detailed illustration of the following components:

- ① ▼ Button ② ▲ Button ③ ⏻ Button

- Charge battery until green LED is lit (rechargeable units only).
- Turn the instrument ON with the ⏻ button. Instrument performs a self test and sensor warm up, then displays sensor info and set points.
- A flow test must be performed during startup. Block pump inlet when instructed.
- Fresh Air Setup: press buttons corresponding with 'YES' or 'NO' on the display.
- To page through the screens, use the ▼ button from the Normal Operation screen, follow display instructions and press corresponding buttons to advance to next page.
- To reset alarms press ▲ button.
- Perform a Bump Test/Calibration Check daily for each installed sensor.
- To calibrate: ensure you are in fresh air and press and hold the ▲ button and follow instructions on the screen. See operating manual for complete calibration instructions.
- To activate/deactivate MotionAlert feature, use the ▼ button until MotionAlert screen and follow display instructions.
- To activate InstantAlert, press and hold ▼ button.
- Turn the instrument OFF by holding the ⏻ button.

WARNING

This quick start card provides only a brief description of the operating instructions for the ALTAIR 5/ALTAIR 5X Multigas Detector. The user of this instrument must be fully aware of the limitations and instructions supplied in the operating instruction manual. Failure to follow the instructions may lead to serious personal injury or death.

ALTAIR5/5X Multigas Detector Quick Start Card

